

4/1/16

REPEAT BIOMETRY IF:

- AXIAL LENGTH LESS THAN 21.2MM OR GREATER THAN 26.6MM
- K READINGS LESS THAN 41D OR GREATER THAN 47D
- AXIAL LENGTH DIFFERENCE OF 0.7MM OR MORE
- DELTA 'K'S GREATER THAN 2.5D
- DIFFERENCE IN MEAN CORNEAL POWER OF > 0.9 DIOPTRES

ALWAYS RECORD REMARKS ON HARD COPY AND MEDISOFT AND NOTE IF MEASUREMENTS HAVE BEEN RECHECKED

THERE IS NO SPECIFIC NEED TO REPEAT BIOMETRY BECAUSE OF THE LENGTH OF TIME FROM THE PREVIOUS TEST UNLESS REQUESTED BY CLINICIAN

IF WORSE EYE REQUIRES OCUSCAN MEASURE BOTH EYES WITH OCUSCAN FOR COMPARISON BUT RECORD BETTER EYE ON MEDISOFT WITH IOL.

PSEUDO PHAKIC EYES DO NOT NEED TO BE MEASURED UNLESS REQUESTED BY CONSULTANT